

PT. B.D.SHARMA PGIMS/UHS ROHTAK.

e-Tenders Portal: www.etenders.hry.nic.in Institute Website: www.uhsr.ac.in

Advt. No.PS/2018/5980

Dt. 21/5/18

CORRIGENDUM

CORRIGENDUM TO GROUP NO.14 REGARDING RATE CONTRACT FOR FIVE YEARS (EXTENDABLE) FOR SUPPLY OF CONSUMABLES/REAGENTS FOR USE IN 01 NO.FULLY AUTOMATED IMMUNO ANALYZER TO BE INSTALLED FREE OF COST (on reagents consumption basis) IN THE DEPTT. OF BIOCHEMISTRY PGIMS ROHTAK AGAINST LONG TERM E-TENDER NOTICE NO.UHSR/PURCHASE/04/2018/15 DT.26.04.2018 DUE ON 29.05.2018 UPTO 5.00 PM AND TO BE OPENED ON 30.05.2018 AT 11.00 AM—PREBID HELD ON 09.05.2018.

FOLLOWING CHANGES / ADDITIONS IN THE SPECIFICATIONS (AFTER PREBID HELD ON 09.05.2018) HAVE BEEN MADE IN THE ABOVE GROUP:


CHANGES:

- Minimum Throughput of 180 tests per hour with random access.
- On board reagent stability should be 14-60 days and calibration of the parameter should be typically lot based. No daily calibration should be required by the system to save the reagent.
- European CE / USFDA approved.

ADDITIONS:

SR.	NAME OF INVESTIGATION	QUANTITY
1.	ACTH	OPTIONAL
2.	Cortisol	
3.	DHEA	
4.	Insulin	

KINDLY FOLLOW e-Tenders Portal: www.etenders.hry.nic.in, Institute website: www.uhsr.ac.in.


21/5/18
Sr. Professor I/c Purchase