

**PT. B.D.SHARMA PGIMS/UHS ROHTAK**  
**e-TENDER NOTICE NO.UHSR/PURCHASE/COVID-19/2020/02 DT. 08.04.2020**  
**(Due on 15.04.2020 upto 5.00 PM and to be opened on 15.04.2020 at 06.00 PM)**  
**e-Tender Portal: [www.etenders.hry.nic.in](http://www.etenders.hry.nic.in), Institute website: [www.uhsr.ac.in](http://www.uhsr.ac.in)**

**CORRIGENDUM**

**E - Tender Group 01**

<b>Sr. No.</b>	<b>Specifications in original E- Tender</b>	<b>Specifications now after corrigendum</b>
6.	It should have in-built programmable ultrasonic nebulizer producing less than 3-micron meter particle size.	It should have in-built programmable ultrasonic/ pneumatic nebulizer producing less than 3-micron meter particle size. (in case of pneumatic nebuliser, certification to be provided)
11.	Should measure following Respiratory Mechanics/ Manoeuvres:  Static Compliance and Resistance, Low Inflection point (LIP) and upper inflection point (UIP), Maximum Inspiratory Pressure, Work of Breathing, Lung recruitment manoeuvre & Monitoring.	Should measure following Respiratory Mechanics/ Manoeuvres:  Static/Dynamic Compliance and Resistance, Low Inflection point (LIP) and upper inflection point (UIP), Maximum Inspiratory Pressure, Work of Breathing (optional), Lung recruitment manoeuvre & Monitoring.

**Sd/-**  
**Professor I/c Purchase**