

PT. B.D.SHARAMA POSTGRADUATE INSTITUTE OF MEDICAL SCIENCES
ROHTAK

E-mail:- quotations.pgims@uhsr.ac.in

Invitation of e-quotations for Fibre Optics Coldlight Source with Mobile Stand

Inquiry No: 01, Department of Pediatrics, PGIMS Rohtak

Inquiry Issue Date: 1-04-2026

Last Date of Submission: 2-04-2026

Rate quotation in sealed envelope/Password protected quotation are hereby invited from the valid firms and suppliers by the undersigned on behalf of the Director, PGIMS, Rohtak for supply of consumables as per Annexure-I for the Institute as per terms & conditions mentioned below. The filled e-quotations must receive through registered e-mail (Password protected) on or before 2-04-2026 at 3.00PM & quotation opened on 2-4-2026 at 3.30PM

Item Name Fibre Optics Coldlight Source with Mobile Stand for the Department of Pediatrics
Closing Date :- 2-4-2026 Time 3.00PM

1. Terms & Conditions:-

- The quotations received sealed/ online after the deadline shall not be entertained under any circumstances whatsoever. In case of any delay this institute will not be responsible.
- The online submission though password protected mode will be restricted**
- Quotations must be in online prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director of their authorized representative. In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
 - Quotations must be submitted in the office of the undersigned before deadline of submitting the quotation.
 - Rates must be quoted in Indian Rupees and as per the format specified taxes extra if any must be written separately.
 - Rates must be quoted F.O.R. basis (including Freight charges).
 - No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
 - The rates quoted must be valid for 90 days minimum or as per State Govt. instructions issued from time to time from the date of opening and silence of any tendered on this issue shall be treated as agreed with this condition.



h. Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified

i. Any conditional quotations shall be rejected summarily.

j. **Delivery period – As per supply order on issuing by the office.**

k. **Liquidated Damages: -** If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5% per week of the total order values shall be levied subject to maximum of 10% of the total order value on which amount when there is no EMD.

l. **Payment terms: -** Payment will be only after satisfactorily complete delivery/ commissioning of material and after inspection by Inspection committee.

m. Pt. BDS PGIMS, Rohtak reserves the right to increase or decrease quantity and / or amount of work. Decision of quantity of material in the Pt. BDS PGIMS, Rohtak will be final in this regard.

n. Pt. BDS PGIMS, Rohtak reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the Pt. BDS PGIMS, Rohtak will be final in this regard.

o. The Sealed Quotation on the envelop due date & product name mandatory. If not mention on the envelop. The quotation will be rejecting.

Special Terms and Conditions:-

1. Bidder must quote the product as per specifications provided in Annexure-1.
2. The supplier may be asked to submit the sample of quoted make for technical evaluation, to the Pt. BDS PGIMS, Rohtak, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.
3. Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then Pt. BDS PGIMS, Rohtak has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained.
4. **Jurisdiction:** The jurisdiction will be district Rohtak only.
5. It is directed to IT Cell that requisition of above item may also be placed on website of Pt. BDS PGIMS, Rohtak for vide publicity to invite a competitive manner of purchase.
6. Samples will be provided by the firm with quotations, otherwise quotation not accepted for the institute.

Committee

Encl: Annexure-I (Specification)

Annexure -II (Format of Quotation)



Annexure-II

Sr. No.	Name of the items	Req. Qty.	Specification	Remarks
1.	Aireobpic COOB light Source with mobile stand	1	Attached as Annexure (On the letter head of the firm)	For Online Tender

Price Bid Form

To

The Director,
Pt. BDS PGIMS, Rohtak.

- I/ We _____ submitted the quotation for enquiry No "quotation for supply _____ against the inquiry No _____ due on dated _____ at Pt. BDS PGIMS, Rohtak.
- I/ We thoroughly examined, understood and accepted terms and conditions given in the enquiry document, failing which my quotation will be rejected our rightly.
- The Vendor should also certify that the rates quoted in the quotation are not more than the MRP of the manufactures.
- I/ We hereby offer to supply at the following rates .

Sr: No.	Particular	Quantity	Quoted If make any	Price/ Unit Exclusive of Tax (INR)	GST/CST/ST
1.					
2.					

Date _____

Place _____

Phone _____

(Signature of Authorized Person)

Name _____



PGIMS, Rohtak

Fiberoptics Cold light source with mobile stand.

Qty: 1

Approx cost: Rs 1,00,000/-

Technical Specifications :-

1. High intensity cold light source for detection of pneumothorax and finding artery in small children easily.
2. It Should be light weight and portable.
3. Should have minimum 1 LED/XENON as source of light.
4. Should have alarm system and indication in case of temperature instability, failure of fan and in dimming of light if any.
5. The fiberoptic cable should be approximately 1 meter in length.
6. The tip of the probe (light source) should not get heated on use.
7. A good quality case for storage of fiberoptic cable and light source.
8. The unit should run on 220-240 volts A.C.
9. Should have easy mobility with availability of stand.
10. Shape and size the probe suitable for use in pediatric and newborns patients.
11. Total Weight of the machine should be between 3 to 5 kg.
12. Should have Class 1 protection against electric shock.
13. Operating temperature of the machine should not exceed 30°C.
14. Should be of minimum ISO Certification product.
15. Warranty of atleast 1 year with onsite repair available and availability of brochure including 24 × 7 toll free helpline number.
16. Atleast 1 extra 3 meter power cord, 1 spare illuminating lamp and 1 extra optical fibre cable should be provided as accessory alongwith the equipment.


Dr. KAPIL BHALLA
Professor
Pediatrics Department
PGIMS, Rohtak


Dr. Vandana Arya
Professor
Pediatrics Department
PGIMS, Rohtak


Dr. Anjali Verma
Professor
Pediatrics Department
PGIMS, Rohtak